

Basic peptide

Peptide with 32 aminoacids and isoelectric point of 11.1

Experimental

Column: SVEA C18 Gold, 5 μm 110 \AA 4.6 x 150 mm
Instrument: HPLC
Mobile phase: A: 0.1% TFA in water; B: 0.085% TFA in acetonitrile.
Gradient: 32-37% B in 40 minutes
Flow rate: 1.0 mL/min
Column temperature: 30°C
Detector: UV 220 nm

