


Hypotensive diuretic

Experimental

Column: SVEA Core C18, 2.6 μm , 4.6 x 100 mm
Instrument: HPLC
Flow rate: 1.0 mL/min
Column temperature: 40°C
Detector: UV 250 nm
Sample:

1. 4-Amino-6-chloro-1,3-benzenedisulfon-amine
2. Hydrochlorothiazide
3. p-Aminoacetophenone

Mobilephase: Acetonitrile/100mM phosphate buffer pH 3.0 (20/80)



Mobilephase: Acetonitrile/0.1% formic acid (20/80)

