

Resorcinol and related compounds

Experimental

Column: SVEA Core C18, 2.6 μm , 4.6 x 100 mm

Instrument: HPLC

Mobile phase: Methanol/0.1% formic acid (10/90)

Flow rate: 1.0 mL/min

Column temperature: 30°C

Detector: UV 280 nm

Sample:

1. Hydroquinone
2. Resorcinol
3. Pyrocatechol
4. Phenol

