

Tocopherols

Experimental

Column: SVEA C18 Gold, 5 µm 110 Å 4.6 x 150 mm

Instrument: HPLC

Mobile phase: Methanol/water (95/05)


Flow rate: 2 mL/min

Column temperature: 30°C

Detector: UV 210 nm

Sample:

1. Delta-tocopherol
2. Gamma-tocopherol
3. Alfa-tocopherol



Performance

$R_s 2,1 = 4.58$

Tailing (2) = 1.07